# VCCM600M



#### MEDICAL AC/DC CONDUCTION COOLED CONFIGURABLE POWER SUPPLY

## DATA SHEET





The medically approved VCCM600M conduction cooled configurable power supply delivers a silent 600 Watts and up to 750 Watts of peak power for 5 seconds in a rugged 4" x 7" package a nd is the ultimate power solution for applications where reliability or audible noise are of concern. The product combines the advantages of a modular and configurable power supply with the high reliability of a fan-less architecture. Depending on your application, the VCCM600M can be configured as a conduction, convection or forced air cooled solution and this versatility allows the unit to be seamlessly integrated across a vast range of applications, which makes it perfect for standardising your power platform.

Designed with highest reliability and versatility in mind, the VCCM600M is suitable for applications ranging from the most controlled to the harshest of environments. Standard features include full output voltage adjust range, externally controllable voltage and current and series & paralleling of outputs. The unique design approach and heat dissipation techniques allows the unit to be mounted in virtually any orientation giving system designers even more flexibility. The series is approved to latest medical safety (IEC/UL60601-1-2 3rd Edition) and EMC standards and features market leading specifications and design in application support.

#### **MAIN FEATURES**

- 600 Watts output (Vin >120V<sub>RMS</sub>)
- Peak power capability (750W 5sec)
- 7" x 4" x 1.61" footprint
- Convection/Conduction/Forced-Air cooled
- Modular & user configurable
- Low power standby mode (<1W)</li>

## APPLICATIONS

- Medical & diagnostic equipment
- Test & Measurement equipment
- Robotics
- Oil & Gas

## CUSTOMER BENEFITS

- Fast time to market
- 24 hrs samples from distribution
- Safety & EMC certified
- World class engineering support

- High efficiency up to 90%
- Additional 5V 1A bias supply
- Remote voltage & current programming
- Current output signal
- Accurate current sharing
- Programmable start-up state (Laser Apps)
- Telecommunications
- Laboratory & Analysis equipment
- Display
- Avionics

- IEC60601 Ed. 3 (Immunity to Ed. 4)
- MIL-STD 810G
- MIL-STD 461F
- MIL-STD 704F
- SEMI F47 compliant
- 5 Year warranty

Lasers

- LED lighting
- High vibration & shock
- Retrofit of legacy PSUs

- Proven technology
- Eliminates custom design costs
- Field replaceable
- Low cost of ownership
- Technology consolidation
- Supplier consolidation

VCCM Medical System Datasheet | DOC6044r00 | Release date 19/10/2016

中国区总代理:上海佳舍珀电子科技有限公司 电话:021-61999334 | 手机:13764303579 | 电子邮件:roman.xiao@fitpower.cn | www.fitpower.cn

## **SPECIFICATIONS**

	INPUT MODULE SPECIFICATIONS				
Parameter	Details	Min	Typical	Max	Units
AC Input Voltage	Nominal range is 100V <sub>RMS</sub> to 240V <sub>RMS</sub>	85		264	V <sub>RMS</sub>
AC Input Frequency	Contact factory for 400Hz operation.	47	50/60	63	Hz
DC Input Voltage		120		370	V <sub>DC</sub>
Output Power Rating	De-rate linearly from 600Watts at 120V $_{\mbox{\tiny RMS}}$ to 425Watts at 85V $_{\mbox{\tiny RMS}}$			600	Watts
Input Current	600Watts output at 120 V <sub>RMS</sub> input			б	Amps
Input Current Limit			7		Amps
Inrush Current	265V <sub>RMS</sub> , 25°C (cold start)			20	Amps
Fusing	Each line fused (5x20 Fast acting)			8	Amps
Efficiency	See graphs			90	%
No load Power consumption	All outputs fitted and disabled/enabled		10/21		Watts
Standby Power	Latched off state, 120V <sub>RMS</sub>		0.5	1	Watts
Power Factor			0.99		
Holdup	600Watts output at 120V <sub>RMS</sub> input	17	20	21	mS
UVP	Turn on under voltage protection	78		84	V <sub>RMS</sub>
Over temperature	Internally monitored.	115		125	°C
Reliability <sup>(1)</sup>	Input module			1.1	FPMH
	Transformer module			0.4	FPMH
Warranty	Standard terms and conditions apply			5	Years
Size	177.8 (L) x 101.6 (W) x 41.0 (H). See diagram for tolerance details		•		mm
Weight	650 + 100 per output module				Grams
Note 1.	30°C base & ambient, 100% load, SR332 Issue 2 Method I, Case 3, Ground, Fixed, Contro	lled			

	GLOBAL SIGNALS SPECIFICATIONS				
Parameter	Details	Min	Typical	Max	Units
Bias Voltage		4.8	5	5.2	Volts
Bias Current				1	Amps
AC_OK Voltage	Low output level/High output level	0/4.8	0.03/5	0.1/5.2	Volts
AC_OK Current				10	mA
Power Good Voltage	Open collector output. Low output level. All slots. Absolute maximum = 6V.	0.1		0.3	Volts
Power Good Current	Open collector output. Current sink only. All Slots.			50	mA
Tsns Voltage	Typical at 0°C internal temperature, 19.5mV/°C	0	0.4	5	Volts
Tsns Current				100	uA
Inhibit Voltage	Low input level/High input level. All slots.	0/2.5		0.8/6	Volts
Inhibit Current	10k input impedance. All slots.			1	mA

	OUTPUT MODULE SPECIFICATION SUMMARY											
MODEL	Out	put Volta	age	Output	Rated	Peak	Load	Line	Cross	Ripple &	FPMH <sup>(1)</sup>	Feature
MODLL	Min.	Nom.	Max.	Current	Power	Power	Reg.	Reg.	Reg.	Noise	1 1 1011 1	Set <sup>(2)</sup>
OPA	1.5V	5V	7.5V	25A	125W	187.5W	±50mV	±5mV	±10mV	50mV <sub>PP</sub>	0.5	ABCDEFG
OPB	4.5V	12V	15V	15A	150W	225W	±100mV	±12mV	±24mV	120mV <sub>PP</sub>	0.5	ABCDEFG
OPC	9V	24V	30V	7.5A	150W	225W	±150mV	±24mV	±48mV	240mV <sub>PP</sub>	0.5	ABCDEFG
OPD	18V	48V	58V	3.75A	150W	217.5W	±300mV	±48mV	±96mV	480mV <sub>PP</sub>	0.5	ABCDEFG
OPE	4.5V	5V	5.5V	100A	500W	TBD	TBD	TBD	TBD	TBD	TBD	AEFGHIJ
OPF	10.8V	12V	13.2V	50A	600W	TBD	TBD	TBD	TBD	TBD	TBD	AEFGHIJ
OPG	21.6V	24V	26.4V	25A	600W	TBD	TBD	TBD	TBD	TBD	TBD	AEFGHIJ
OPH	43.2V	48V	52.8V	12.5A	600W	TBD	TBD	TBD	TBD	TBD	TBD	AEFGHIJ
Note 1.	Output r	nodule, 30°	C base, 10	0% load, SR332	issue 2 Method	d I, Case 3, Gro	und, Fixed, Co	ontrolled				

 Note 1.
 Output module, 30°C base, 100% load, SR332 issue 2 Method I, Case 3, Ground, Fixed, Controlled

 Note 2.
 A = Remote Sense, B = External Voltage control, C = External constant current control, D = Current output signal, E = Current share, F =Over Voltage protection, G = Over temperature protection, H = BF rating, I = PMBUS capability, J =Internal ORing mosfet

SAFETY SPECIFICATIONS Parameter Details Max Notes Input to Output (2 MOPP) 4000  $V_{\text{AC}}$ Input to J2 standby control (2 MOPP) 4000 V<sub>AC</sub> Input to Chassis (1 MOPP) 1500  $V_{\text{AC}}$ Isolation Voltages  $V_{\text{DC}}$ Global signals (J3) to Output/Chassis 500 Output to Output/Chassis (Standard modules) 500  $V_{\text{DC}}$ Output to Output/Chassis (BF Rated modules, 1 MOPP) 1500 VAC Normal condition, 264Vac, 63Hz, 25°C Earth Leakage Current 200 uA Standard modules NC/SFC 20/200 uA Touch Leakage Current BF rated modules NC/SFC uА TBD/TBD Standard modules 264Vac, 63Hz, 25°C NC/SFC Not applicable uA Patient Leakage Current BF rated modules 264Vac, 63Hz, 25°C NC/SFC TBD/TBD uA

	INSTALLATION SPECIFICATIONS									
Parameter Details Parameter Details										
Equipment class		Flammability Rating	94V-2							
Overvoltage category	II	Ingress protection rating	IP10							
Material Group	IIIb (indoor use only)	ROHS compliance	2011/65/EU							
Pollution degree	2	Intended usage environment	Home Healthcare							

VCCM Medical System Datasheet | DOC6044r00 | Release date 19/10/2016

Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 459 1161 | www.vox-power.com

中国区总代理:上海佳舍珀电子科技有限公司 电话:021-61999334 | 手机:13764303579 | 电子邮件:roman.xiao@fitpower.cn | www.fitpower.cn

	ENVIRONMENTAL SPECIFICA	<b>FIONS</b>				
Parameter	Details	Non-Op	erational	Opera	- Units	
Parameter	Details	Min	Max	Min	Max	- Units
Air Temperature	Operational limits subject to appropriate de-ratings	-51	+85	-40(1)	70	°C
Humidity	Relative, non-condensing	5	95	5	95	%
Altitude		-200	5000	-200	3000	m
Shock	EN 60068-2-27: Half sine, 3 axes, 3 positive & 3 negative. 810G: Method 516.6, Procedure IV, Transit drop		50, 11		30,18	g, mS
Vibration	EN 60068-2-6: Sine, 10 – 500 Hz, 3 axes, 1 oct/min., 10 cycles each axis EN 60068-2-64: Random, 5 – 500 Hz, 3 axes, 30 min. 810G: Method 514.6, Procedure I (General Vibration) Category 4 (Trucks & Trailers, Composite wheeled vehicle), Figure 514.6C-3. Category 7 (Aircraft, Jet cargo), Figure 514.6C-5 General exposure Category 24, (All, Minimum integrity) Figure 514.6E-1		0.02,2.56		2 0.0122,1	g g²/Hz, g <sub>RMS</sub>
Thermal shock	MIL-STD-810G Method 503.5 Procedure I-C. Multi-cycle. 3 shocks.	-51	85			°C

Notes 1. Some specifications may not be met below -20°C

ELECTROMAGNETIC COMPLIANCE – EMISSIONS							
Phenomenon	Basic EMC Standard	Test Details					
Radiated emissions, electric field	EN55011/22	Class B compliant					
Radiated emissions, electric field, 30Hz-18GHz.	MIL-STD-461F: RE102 (Ground, Fixed)	Compliant (When mounted in enclosure)					
Conducted emissions	EN55011/22, FCC part 15, CISPR 22/11	Class B compliant					
Conducted emissions, power leads, 10kHz-10Mhz.	MIL-STD-461F: CE102	Compliant					
Harmonic Distortion	IEC61000-3-2	Compliant					
Flicker & Fluctuation	IEC61000-3-3	Compliant					

ELE	CTROMAGNETIC COMPLIA	NCE – IMMUNITY
Phenomenon	Basic EMC Standard	Test Details
Electrostatic discharge	IEC61000-4-2	Test level 4: 15kV air, 8kV contact, IEC60601-1-2:2014 compliant
Radiated RF EM fields	IEC61000-4-3	Test Level 3: (10V/m, 80MHz-2.7GHz) sine wave AM 80% 1kHz
Proximity fields from RF wireless communications equipment	IEC61000-4-3	Test levels as per IEC60601-1-2:2014 Table 9
Radiated susceptibility, electric field, 2 MHz to 40 GHz.	MIL-STD-461F: RS103	20V
Electrical Fast Transients/bursts	IEC61000-4-4	Test Level 3: (2kV Power, 1kV I/O) 5kHz(ed3) & 100kHz(ed4)
Conducted susceptibility, Bulk cable injection, impulse excitation	MIL-STD-461F: CS115	
Surges	IEC61000-4-5	Test Level 3: 1kV L-N, 2kV L-E. As per IEC60601-1-2:2014
Conducted susceptibility, damped sinusoidal transients, cables and power leads, 10kHz-100MHz	MIL-STD-461F: CS116	
Shipboard Electric Power. Voltage Spike Test	MIL-STD-1399, SECTION 300A	Type 1, 115V 60Hz single phase
Conducted disturbances induced by RF fields	IEC61000-4-6	Test Level 3: 10V, 0.15 to 80Mhz sine wave AM 80% 1kHz
Conducted susceptibility, power leads, 30Hz-150kHz	MIL-STD-461F: CS101	
Conducted susceptibility, Bulk cable injection, 10kHz- 200Mhz	MIL-STD-461F: CS114	
Power Frequency Magnetic Fields	IEC61000-4-8	Test level 4: 30A/m 50Hz
Radiated susceptibility, Magnetic field, 30Hz-100kHz	MIL-STD-461F: RS101	
Voltage Dips	IEC61000-4-11 <sup>(2)</sup>	0% 10ms, 0% 20ms, 70% 0.5s (Criterion A) 40% 200mS (Criterion A at 240V and Criterion C at 100V)
Voltage Sag Immunity	SEMI-F47-0706 <sup>(2)</sup>	0% 20mS, 70% 0.5s, 80% 1s,80% 10s,90% continuous (Criterion A) 50% 200mS (Criterion A at 240V and Criterion C at 100V) Criterion A is achieved for full power when Vin >=160V Criterion A is achieved at all input voltages when Pout <= 350W
Voltage interruptions	IEC61000-4-11	0% 250/300 cycle as per IEC60601-1-2:2014 (Criterion C)
Aircraft Electric Power Characteristic	MIL-STD-704F	SAC102,104,105,109,110 (MIL-HDBK-704-2) & SXF102,104,105,109,110 (MIL-HDBK-704-6)
Notes		, , , , , , , , , , , , , , , , , , ,


Notes:

1. Criterion A = No degradation of performance or loss of function.

Criterion B = No degradation of performance or loss of function is allowed below a performance level specified by the manufacturer. Criterion C = Temporary loss of function is allowed, provided the function is self-recoverable. Tested at nominal range (100V to 240V). Line deratings applied where appropriate.


AGENCY APPROVALS						
Standard	Details	File				
IEC 60601-1:2005/AMD1:2012/COR1:2014	3rd Edition	UL: E316486				
UL60601-1:2006						
CAN/CSA - C22.2 No. 60601- 1:14 - Edition 3	Medical Equipment Part 1: General requirements for basic Safety and essential Performance					
ANSI/AAMI ES60601-1(2005 +C1:09 +A2:10)	Medical Equipment Part 1: General requirements for basic Safety and essential Performance					
CE MARK	LVD 2014/35/EU, EMC 2014/30/EU					
CB certificate and report available on request						

VCCM Medical System Datasheet | DOC6044r00 | Release date 19/10/2016




VCCM Medical System Datasheet | DOC6044r00 | Release date 19/10/2016

Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 459 1161 | www.vox-power.com 中国区总代理:上海佳舍珀电子科技有限公司 电话:021-61999334 | 手机:13764303579 | 电子邮件:roman.xiao@fitpower.cn | www.fitpower.cn



VCCM Medical System Datasheet | DOC6044r00 | Release date 19/10/2016

	PINOUTS
Circuit	Details
	J1 – Mains Input
1	Live
2	Neutral
3	Earth
	J2 – Standby control
1	Standby control negative
2	Standby control positive
	J3 – Global Signals
1	Slot 4 - Power Good
2	Slot 4 - Inhibit
3	Slot 3 - Power Good
4	Slot 3 - Inhibit
5	Slot 2 - Power Good
6	Slot 2 - Inhibit
7	Slot 1 - Power Good
8	Slot 1 - Inhibit
9	Temperature sense (T <sub>SNS</sub> )
10	AC OK
11	+5V (Bias Supply 1A)
12	COM
	J4 -Output Signals
1	- Sense
2	+ Sense
3	COM
4	l Control
5	V Control
6	+5V (Bias Supply 20mA)



Unless stated otherwise, All dimensions are in milimeters and in accordance with DIN2768-1/-2 CLASS C

	MATING CONNECTORS			
Ref.	Details	Manufacturer	Housing	Terminal
J1 - Mains Input	3 Pin, Barrier, 6-32 Steel Screws, 0.8 Nm or 7 Lb-In Torque <sup>(1)</sup>			
J2 - Standby control	2 Pin, 1.25mm, with Friction Lock, 28-30AWG	MOLEX	0510210200	0500588000
J3 - Global Signals	12 Pin, 2mm, with Friction Lock, 24-30 AWG, WIRE TO BOARD	MOLEX	0511101260	0503948051
JS - GIODAI SIGNAIS	12 Pin, 2mm, with Friction Lock, 24-30 AWG, IDT CABLE TO BOARD	MOLEX	0875681273	
J5 - Output Signals	6 PIN, 1.25mm, with Friction Lock, 28-30AWG	MOLEX	0510210600	0500588000
Output Power	Positive/Negative, M4 terminal, use appropriately rated crimp terminal			
2. Direct equiv	AWG, 300V, 16A, 105°C, use appropriately rated crimp terminal. /alents may be used for any connector parts. ust be rated 105°C min, equivalent to UL1015	·		

#### PART NUMBERS AND ORDERING INFORMATION

Series name		VCCM600	Μ	-	А	A	В	В	-	0	0	0	 Factory Use
Leakage Current Medical	] 												Use '0' for unused slots.
Slot 1 - Output #	<u> </u>												 Slot 4 - Output #
Slot 2 - Output #	<u> </u>												Slot 3 - Output #

All specifications are believed to be correct at time of publishing. Vox Power Ltd reserves the right to make changes to any of its products and to change or improve any part of the specification, electrical or mechanical design or manufacturing process without notice. Vox Power Ltd does not assume any liability arising out of the use or application of any of its products and of any information to the maximum extent permitted by law. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any products of Vox Power Ltd. VOX POWER LTD DISCLMIS ALL WARRANTES AND REPRESENTATIONS OF ANY KIND WHETHER EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, IMPLIED WARRANTES OF SUITABILITY, FINISES FOR PURPOSE, MERCHANTABILITY AND NONINFRINGEMENT. Please consult your local distributor or Vox Power directly to ensure that you have the latest revision before using the product and refer to the latest relevant user manual for further information relating to the use of the product. Vox Power Ltd products are not intended for use in connection with life support systems, human implantations, nuclear facilities or systems, aircraft, spacecraft, military or naval missile, ground support or control equipment used for the purpose of guidance navigation or direction of any aircraft, spacecraft or any other application where product failure could lead to loss of life or catastrophic property damage. The user will hold Vox Power Ltd harmless from any loss, cost or damage resulting from its breach of these provisions.

VCCM Medical System Datasheet | DOC6044r00 | Release date 19/10/2016

Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 459 1161 | www.vox-power.com

中国区总代理:上海佳舍珀电子科技有限公司 电话:021-61999334 | 手机:13764303579 | 电子邮件: roman.xiao@fitpower.cn | www.fitpower.cn